

Certificado No. (Certificate #): 53418

Equipo (Instrument): Calibrador Pie de Rey

Fecha de Calibración (Calibration Date): 2025-01-29

Marca (Brand): Performance Tool

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	1 mm	1.02 mm	20 µm	± 30 µm	± 2.0E+02 µm
Exterior / (Exterior)	10 mm	10.11 mm	110 µm	± 30 µm	± 2.0E+02 µm
Exterior / (Exterior)	50 mm	50.13 mm	130 µm	± 30 µm	± 2.0E+02 µm
Exterior / (Exterior)	100 mm	100.18 mm	180 µm	± 30 µm	± 2.0E+02 µm
Exterior / (Exterior)	150 mm	150.21 mm	210 µm	± 30 µm	± 2.0E+02 µm
Interior / (Interior)	20 mm	19.90 mm	-97 µm	± 30 µm	± 8.2E+00 µm
Interior / (Interior)	100 mm	99.96 mm	-43 µm	± 30 µm	± 8.2E+00 µm
Profundidad / (Depth)	30 mm	30.05 mm	53 µm	± 30 µm	± 7.1E+00 µm
Profundidad / (Depth)	50 mm	50.05 mm	50 µm	± 30 µm	± 7.1E+00 µm