

Certificado No. (Certificate #):		53418		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-01-29		Marca (Brand):		Performance Tool	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		1 mm	1.02 mm	20 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		10 mm	10.11 mm	110 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		50 mm	50.13 mm	130 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		100 mm	100.18 mm	180 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		150 mm	150.21 mm	210 µm	± 30 µm	± 7.4E+00 µm	
Interior / (Interior)		20 mm	19.90 mm	-100 µm	± 30 µm	± 8.2E+00 µm	
Interior / (Interior)		100 mm	99.90 mm	-100 µm	± 30 µm	± 8.2E+00 µm	
Profundidad / (Depth)		30 mm	30.05 mm	50 µm	± 30 µm	± 7.1E+00 µm	
Profundidad / (Depth)		50 mm	50.05 mm	50 µm	± 30 µm	± 7.1E+00 µm	