

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	29.9 °C	-0.162 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.0 %HR	0.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.652 %HR	48.0 %HR	-2.652 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	71.0 %HR	0.552 %HR	± 5.0 %HR	± 1.9E+00 %HR