Patrón (Standard)		UBP (UUT)		Error (Error) E.M.P (Tolerance)		Incertidum	Incertidumbre (Uncertainty)	
1.0625	in	1.062	in	-0.0005 in	$\pm 0.0002 in$	±	5.8E-04 in	
1.1	in	1.100	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.125	in	1.125	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.2	in	1.200	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.25	in	1.250	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.3	in	1.300	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.40	in	1.400	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.5	in	1.500	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.600	in	1.600	in	0 in	± 0.0002 in	±	5.8E-04 in	
1.80	in	1.800	in	0 in	± 0.0002 in	±	5.8E-04 in	
2.0	in	2.000	in	0 in	± 0.0002 in	±	5.8E-04 in	