

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.00 mm	0.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm
1.25 mm	1.24 mm	-10 µm	± 30 µm	±	5.8E+00 µm
1.50 mm	1.49 mm	-10 µm	± 30 µm	±	5.8E+00 µm
2.00 mm	1.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm
3.00 mm	2.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm
5.00 mm	4.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm
6.00 mm	5.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm
8.00 mm	7.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm
10.00 mm	9.99 mm	-10 µm	± 30 µm	±	5.8E+00 µm