Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	4.00 in	4.001 in	0.001 in	± 0.0028 in	± 1.3E-03 in
Exterior / (Exterior)	4.80 in	4.801 in	0.001 in	± 0.0028 in	± 1.3E-03 in
Exterior / (Exterior)	6.40 in	6.401 in	0.001 in	± 0.0028 in	\pm 1.3E-03 in
Exterior / (Exterior)	7.20 in	7.201 in	0.001 in	± 0.0028 in	\pm 1.3E-03 in
Exterior / (Exterior)	8.00 in	8.001 in	0.001 in	± 0.0031 in	\pm 1.3E-03 in
Interior / (Interior)	0.80 in	0.800 in	0.000 in	± 0.0020 in	\pm 7.4E-04 in
Interior / (Interior)	7.20 in	7.200 in	0.000 in	± 0.0028 in	\pm 7.4E-04 in
Profundidad / (Depth)	0.80 in	0.800 in	0.000 in	± 0.0020 in	\pm 8.7E-04 in
Profundidad / (Depth)	7.20 in	7.200 in	0.000 in	± 0.0028 in	± 8.7E-04 in