Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	4.00 in	4.003 in	0.003 in	$\pm 0.0028$ in	± 1.3E-03 in
Exterior / (Exterior)	4.80 in	4.806 in	0.006 in	$\pm 0.0028$ in	± 1.3E-03 in
Exterior / (Exterior)	6.40 in	6.406 in	0.006 in	$\pm 0.0028$ in	± 1.3E-03 in
Exterior / (Exterior)	7.20 in	7.206 in	0.006 in	$\pm 0.0028$ in	± 1.3E-03 in
Exterior / (Exterior)	8.00 in	8.006 in	0.006 in	$\pm 0.0031$ in	± 1.3E-03 in
Interior / (Interior)	0.80 in	0.800 in	0.000 in	$\pm 0.0020$ in	± 7.4E-04 in
Interior / (Interior)	7.20 in	7.204 in	0.004 in	$\pm 0.0028$ in	± 7.4E-04 in
Profundidad / (Depth)	0.80 in	0.803 in	0.003 in	$\pm 0.0020$ in	± 8.7E-04 in
Profundidad / (Depth)	7.20 in	7.202 in	0.002 in	$\pm 0.0028$ in	$\pm$ 8.7E-04 in