

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.021 °C	14.5 °C	-0.521 °C	± 1.0 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.3 °C	-0.749 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	39.0 %HR	3.317 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.846 %HR	72.0 %HR	1.154 %HR	± 5.0 %HR	± 1.9E+00 %HR