

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.05	0.04952	0.480	± 0.6	± 4.3E-01 mm					
0 a 0.1	0.09955	0.450	± 0.6	± 4.3E-01 mm					
0 a 0.15	0.14952	0.480	± 0.6	± 4.3E-01 mm					
0 a 1	0.99946	0.540	± 1	± 4.3E-01 mm					
0 a 2	1.99968	0.320	± 1.4	± 4.8E-01 mm					
0 a 3	2.99961	0.390	± 1.8	± 4.8E-01 mm					
0 a 4	3.99983	0.170	± 2.2	± 4.8E-01 mm					
0 a 5	4.99958	0.420	± 2.6	± 4.8E-01 mm					
0 a 6	5.99969	0.310	± 3	± 3.6E-01 mm					
0 a 7	6.99963	0.370	± 3.4	± 3.6E-01 mm					
0 a 8	7.99978	0.220	± 3.8	± 3.6E-01 mm					
0 a 9	8.99988	0.120	± 4.2	± 3.6E-01 mm					
0 a 10	9.99998	0.020	± 4.6	± 3.6E-01 mm					
0 a 11	10.99976	0.240	± 5	± 3.6E-01 mm					
0 a 12	12.00000	0.000	± 5.4	± 3.6E-01 mm					
0 a 13	13.00002	-0.020	± 5.8	± 3.6E-01 mm					
0 a 14	14.00013	-0.130	± 6.2	± 3.6E-01 mm					
0 a 15	14.99997	0.030	± 6.6	± 3.6E-01 mm					