

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.091 °C	15.1 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.9 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	32.3 °C	0.232 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.202 %HR	21.7 %HR	1.498 %HR	± 1.0 %HR	± 1.9E+00 %HR
50 %HR	50.182 %HR	53.4 %HR	3.218 %HR	± 1.0 %HR	± 1.9E+00 %HR
70 %HR	70.294 %HR	73.6 %HR	3.306 %HR	± 1.0 %HR	± 1.9E+00 %HR