

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.091 °C	15.0 °C	-0.091 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.9 °C	-0.149 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.068 °C	31.9 °C	-0.168 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.202 %HR	23.5 %HR	3.298 %HR	± 1.0 %HR	± 1.9E+00 %HR
50 %HR	50.182 %HR	53.9 %HR	3.718 %HR	± 1.0 %HR	± 1.9E+00 %HR
70 %HR	70.294 %HR	73.7 %HR	3.406 %HR	± 1.0 %HR	± 1.9E+00 %HR