

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.131 °C	15.3 °C	0.169 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.7 °C	0.451 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	33.171 °C	33.5 °C	0.329 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	17.932 %HR	21.7 %HR	3.768 %HR	± 1.0 %HR	± 1.9E+00 %HR
50 %HR	49.169 %HR	50.7 %HR	1.531 %HR	± 1.0 %HR	± 1.9E+00 %HR
70 %HR	73.004 %HR	74.8 %HR	1.796 %HR	± 1.0 %HR	± 1.9E+00 %HR