

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.131 °C	15.3 °C	0.169 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.249 °C	25.7 °C	0.451 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	33.171 °C	33.5 °C	0.329 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	17.932 %HR	18.7 %HR	0.768 %HR	± 1.0 %HR	± 1.9E+00 %HR
50 %HR	49.169 %HR	49.9 %HR	0.731 %HR	± 1.0 %HR	± 1.9E+00 %HR
70 %HR	73.004 %HR	73.5 %HR	0.496 %HR	± 1.0 %HR	± 1.9E+00 %HR