

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.862 °C	30.2 °C	0.338 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	28.8 %HR	-1.770 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.478 %HR	47.6 %HR	-2.878 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	65.7 %HR	-4.748 %HR	± 3.0 %HR	± 1.9E+00 %HR