

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.173 °C	30.6 °C	0.427 °C	± 0.6 °C	± 4.0E-01 °C
30 %HR	28.873 %HR	29.4 %HR	0.527 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.664 %HR	50.1 %HR	-1.564 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	68.5 %HR	-3.696 %HR	± 3.0 %HR	± 1.9E+00 %HR