

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-24.9 °C	0.189 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.051 °C	0.1 °C	0.151 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.274 °C	5.3 °C	0.026 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.087 °C	9.9 °C	-0.187 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.028 °C	24.9 °C	-0.128 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.032 °C	29.9 °C	-0.132 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.100 °C	40.0 °C	-0.100 °C	± 0.45 °C	± 4.0E-01 °C

