

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.570 %HR	29.8 %HR	-0.770 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.488 %HR	50.3 %HR	-0.188 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	69.3 %HR	-1.045 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	90.100 %HR	90.3 %HR	0.200 %HR	± 5 %HR	± 1.9E+00 %HR

