

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-25.4 °C	-0.311 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	-0.522 °C	-0.6 °C	-0.078 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.194 °C	4.9 °C	-0.294 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.509 °C	10.5 °C	-0.009 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.1 °C	0.072 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	30.183 °C	30.2 °C	0.017 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	40.541 °C	40.7 °C	0.159 °C	± 0.45 °C	± 4.0E-01 °C

