

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.570 %HR	30.2 %HR	-0.370 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.488 %HR	51.0 %HR	0.512 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	70.345 %HR	69.4 %HR	-0.945 %HR	± 3 %HR	± 1.9E+00 %HR
90 %HR	90.100 %HR	90.7 %HR	0.600 %HR	± 5 %HR	± 1.9E+00 %HR

