

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.089 °C	-25.4 °C	-0.311 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.051 °C	0.4 °C	0.451 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.274 °C	5.0 °C	-0.274 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.087 °C	9.6 °C	-0.487 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.028 °C	24.8 °C	-0.228 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.032 °C	30.0 °C	-0.032 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.100 °C	39.9 °C	-0.200 °C	± 0.5 °C	± 4.0E-01 °C

