

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.621 °C	0.2 °C	-0.421 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.035 °C	19.8 °C	-0.235 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.277 °C	35.4 °C	0.123 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	32.5 %HR	1.930 %HR	± 2.0 %HR	± 1.9E+00 %HR
55 %HR	50.816 %HR	51.0 %HR	0.184 %HR	± 2.0 %HR	± 1.9E+00 %HR
80 %HR	80.673 %HR	80.8 %HR	0.127 %HR	± 2.0 %HR	± 1.9E+00 %HR