

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1 mm	0.99 mm	-10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	2 mm	1.49 mm	-10 µm	± 30 µm	± 5.8E+00 µm
Profundidad / (Depth)	2 mm	1.99 mm	-10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	3 mm	3.00 mm	0 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	5 mm	4.99 mm	-10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	10 mm	9.99 mm	-10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	15 mm	14.99 mm	-10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	20 mm	19.99 mm	-10 µm	± 30 µm	± 6.1E+00 µm
Profundidad / (Depth)	25 mm	24.99 mm	-10 µm	± 30 µm	± 6.1E+00 µm