

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
5 °C	5.194 °C	5.5 °C	0.306 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	14.991 °C	14.8 °C	-0.191 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.062 °C	30.0 °C	-0.062 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.345 %HR	69.3 %HR	-1.045 %HR	± 3.0 %HR	± 1.9E+00 %HR