

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
299.8 RPM	300 RPM	0.2 RPM	± 2 RPM	±	8.2E-02 RPM
698.8 RPM	700 RPM	1.2 RPM	± 2 RPM	±	1.1E-01 RPM
999.2 RPM	1000 RPM	0.8 RPM	± 2 RPM	±	1.1E-01 RPM
1499.5 RPM	1500 RPM	0.5 RPM	± 2 RPM	±	1.1E-01 RPM
1798.5 RPM	1800 RPM	1.5 RPM	± 2 RPM	±	1.1E-01 RPM