

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.168 °C	10.3 °C	0.132 °C	± 0.3 °C	± 4.0E-01 °C
25 °C	25.219 °C	25.3 °C	0.081 °C	± 0.3 °C	± 4.0E-01 °C
40 °C	40.050 °C	40.1 °C	0.050 °C	± 0.3 °C	± 4.0E-01 °C
30 %HR	30.918 %HR	37.9 %HR	6.982 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.664 %HR	57.7 %HR	6.036 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	79.353 %HR	84.4 %HR	5.047 %HR	± 5.0 %HR	± 1.9E+00 %HR