

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1 mm	1.02 mm	20 µm	± 40 µm	±	5.8E+00 µm
1 mm	1.24 mm	-10 µm	± 40 µm	±	5.8E+00 µm
2 mm	1.50 mm	0 µm	± 40 µm	±	5.8E+00 µm
5 mm	5.00 mm	0 µm	± 40 µm	±	5.8E+00 µm
10 mm	10.01 mm	10 µm	± 40 µm	±	5.8E+00 µm
15 mm	15.01 mm	10 µm	± 40 µm	±	5.8E+00 µm
20 mm	20.02 mm	20 µm	± 40 µm	±	5.8E+00 µm
25 mm	25.00 mm	0 µm	± 40 µm	±	5.8E+00 µm
50 mm	50.00 mm	0 µm	± 40 µm	±	5.8E+00 µm