

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
2 °C	2.085 °C	2.1 °C	0.015 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.152 °C	8.3 °C	0.148 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.539 %HR	51.2 %HR	0.661 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	71.8 %HR	1.557 %HR	± 3.0 %HR	± 1.9E+00 %HR