

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.085 °C	2.2 °C	0.115 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.152 °C	8.4 °C	0.248 °C	± 0.5 °C	± 4.0E-01 °C
50 %HR	50.539 %HR	51.9 %HR	1.361 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.5 %HR	-1.743 %HR	± 3.0 %HR	± 1.9E+00 %HR