

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 2 °C                       | 2.085 °C                   | 2.3 °C             | 0.215 °C               | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 8 °C                       | 8.152 °C                   | 7.9 °C             | -0.252 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 50 %HR                     | 50.539 %HR                 | 51.6 %HR           | 1.061 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.243 %HR                 | 70.7 %HR           | 0.457 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |