

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.994 °C	30.3 °C	0.306 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.394 %HR	31.9 %HR	2.506 %HR	± 3 %HR	± 7.4E-02 %HR
50 %HR	50.509 %HR	53.0 %HR	2.491 %HR	± 3 %HR	± 7.4E-02 %HR
70 %HR	68.934 %HR	70.4 %HR	1.466 %HR	± 3 %HR	± 7.4E-02 %HR