

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.233 °C	29.8 °C	-0.433 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.873 %HR	31 %HR	2.127 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	51.664 %HR	52 %HR	0.336 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	71 %HR	-1.196 %HR	± 5 %HR	± 1.9E+00 %HR