

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.233 °C	30.4 °C	0.167 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.873 %HR	33 %HR	4.127 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	51.664 %HR	55 %HR	3.336 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	75 %HR	2.804 %HR	± 5 %HR	± 1.9E+00 %HR