

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C / 86 °F	30.223 °C / 86.4 °F	30.5 °C / 86.9 °F	0.277 °C / 0.5 °F	± 1.0 °C / 2.0 °F	± 4.0E-01 °C
30 %HR	28.873 %HR	30 %HR	1.127 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	51.664 %HR	50 %HR	-1.664 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	71 %HR	-1.196 %HR	± 5 %HR	± 1.9E+00 %HR