

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.381 °C	0.5 °C	0.119 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.737 °C	20.8 °C	0.063 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.391 °C	40.3 °C	-0.091 °C	± 0.5 °C	± 4.0E-01 °C