

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	15.1 °C	-0.021 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.018 °C	25.1 °C	0.082 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.369 °C	32.1 °C	-0.269 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	20.4 %HR	0.055 %HR	± 1.0 %HR	± 1.9E+00 %HR
50 %HR	50.478 %HR	50.0 %HR	-0.478 %HR	± 1.0 %HR	± 1.9E+00 %HR
70 %HR	70.284 %HR	70.8 %HR	0.516 %HR	± 1.0 %HR	± 1.9E+00 %HR