

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
0 °C	-0.141 °C	0.1 °C	0.241 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.191 °C	15.3 °C	0.109 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.022 °C	30.0 °C	-0.022 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.873 %HR	31.6 %HR	2.727 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.664 %HR	54.8 %HR	3.136 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	72.9 %HR	0.704 %HR	± 3.0 %HR	± 1.9E+00 %HR