

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.141 °C	0.1 °C	0.241 °C	± 0.5 °C	± 4.0E-01 °C
15 °C	15.191 °C	15.2 °C	0.009 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.022 °C	29.9 °C	-0.122 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	28.873 %HR	31.1 %HR	2.227 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	51.664 %HR	50.5 %HR	-1.164 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.196 %HR	71.2 %HR	-0.996 %HR	± 3.0 %HR	± 1.9E+00 %HR