

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.089 °C        | -25.4 °C  | -0.311 °C     | ± 0.95 °C            | ± 4.0E-01 °C                |
| 0 °C              | -0.522 °C         | -0.8 °C   | -0.278 °C     | ± 0.45 °C            | ± 4.0E-01 °C                |
| 5 °C              | 5.194 °C          | 5.2 °C    | 0.006 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 10 °C             | 10.509 °C         | 10.2 °C   | -0.309 °C     | ± 0.45 °C            | ± 4.0E-01 °C                |
| 25 °C             | 25.028 °C         | 25.1 °C   | 0.072 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 30 °C             | 30.183 °C         | 30.2 °C   | 0.017 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |
| 40 °C             | 40.541 °C         | 40.7 °C   | 0.159 °C      | ± 0.45 °C            | ± 4.0E-01 °C                |

