

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
23.3 μm	25.0 μm	1.70 μm	± 0.4 μm	±	5.8E-01 μm
51.6 μm	52.0 μm	0.40 μm	± 0.4 μm	±	5.8E-01 μm
124.6 μm	126.0 μm	1.40 μm	± 0.4 μm	±	5.8E-01 μm
250.3 μm	251.0 μm	0.70 μm	± 7.910 μm	±	5.9E-01 μm
486.00 μm	487.0 μm	1.00 μm	± 14.98 μm	±	6.4E-01 μm
963.00 μm	963.0 μm	0.00 μm	± 29.3 μm	±	7.9E-01 μm