

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.127 °C	10.3 °C	0.173 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.238 °C	31.9 °C	-0.338 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	41.439 %HR	46.3 %HR	4.861 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	75.263 %HR	78.2 %HR	2.937 %HR	± 5.0 %HR	± 1.9E+00 %HR