

Nominal <i>(Nominal)</i>	Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>
0 °C	-0.522 °C	-0.8 °C	-0.278 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.007 °C	9.9 °C	-0.107 °C	± 0.5 °C	± 4.0E-01 °C