

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
1 mm	1.1 mm	100 μm	± 0.002 μm	±	5.8E+01 μm
2 mm	2.0 mm	0 μm	± 0.002 μm	±	5.8E+01 μm
5 mm	5.0 mm	0 μm	± 0.002 μm	±	5.8E+01 μm
10 mm	10.0 mm	0 μm	± 0.002 μm	±	5.8E+01 μm
20 mm	20.0 mm	0 μm	± 0.002 μm	±	5.8E+01 μm
50 mm	50.1 mm	100 μm	± 0.002 μm	±	5.8E+01 μm
100 mm	100.1 mm	100 μm	± 0.002 μm	±	5.8E+01 μm
150 mm	150.2 mm	200 μm	± 0.002 μm	±	5.8E+01 μm
200 mm	200.5 mm	500 μm	± 0.002 μm	±	5.8E+01 μm