

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
10.0 μm	10.2 μm	0.2 μm	± 2.5 μm	±	5.8E-02 μm
15.0 μm	15.3 μm	0.3 μm	± 2.5 μm	±	5.8E-02 μm
23.0 μm	23.3 μm	0.3 μm	± 2.5 μm	±	5.8E-02 μm
52.0 μm	52.3 μm	0.3 μm	± 2.5 μm	±	5.8E-02 μm
155.0 μm	156 μm	1.0 μm	± 4.7 μm	±	6.2E-02 μm
257.0 μm	256 μm	-1.0 μm	± 7.7 μm	±	7.0E-02 μm
511.0 μm	514 μm	3.0 μm	± 15 μm	±	9.7E-02 μm