

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.183 °C	29.8 °C	-0.383 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	31.3 %HR	1.037 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.744 %HR	53.0 %HR	2.256 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	70.2 %HR	0.161 %HR	± 3.0 %HR	± 1.9E+00 %HR