

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 30 °C             | 30.183 °C         | 30.2 °C   | 0.017 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 30 %HR            | 30.263 %HR        | 32.8 %HR  | 2.537 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 50 %HR            | 50.744 %HR        | 53.2 %HR  | 2.456 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |
| 70 %HR            | 70.039 %HR        | 71.8 %HR  | 1.761 %HR     | ± 3.0 %HR            | ± 1.9E+00 %HR               |