

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.183 °C	30.2 °C	0.017 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	31.3 %HR	1.037 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.744 %HR	53.1 %HR	2.356 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	70.3 %HR	0.261 %HR	± 3.0 %HR	± 1.9E+00 %HR