

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.183 °C	28.0 °C	-2.183 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	28.0 %HR	-2.263 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.744 %HR	50.0 %HR	-0.744 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.039 %HR	71.0 %HR	0.961 %HR	± 5.0 %HR	± 1.9E+00 %HR