

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.381 °C	0.6 °C	0.219 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.737 °C	21.0 °C	0.263 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.176 °C	35.6 °C	0.424 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	30.2 %HR	-0.268 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	50.120 %HR	52.1 %HR	1.980 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	70.448 %HR	68.7 %HR	-1.748 %HR	± 3.0 %HR	± 1.9E+00 %HR