

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.381 °C	0.2 °C	-0.181 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.737 °C	20.9 °C	0.163 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.176 °C	35.4 °C	0.224 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.468 %HR	30.0 %HR	-0.468 %HR	± 3.0 %HR	± 1.9E+00 %HR
55 %HR	50.120 %HR	50.6 %HR	0.480 %HR	± 3.0 %HR	± 1.9E+00 %HR
80 %HR	70.448 %HR	70.2 %HR	-0.248 %HR	± 3.0 %HR	± 1.9E+00 %HR