

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 30 °C                      | 30.363 °C                  | 30.8 °C            | 0.437 °C               | ± 1.0 °C                      | ± 4.0E-01 °C                         |
| 30 %HR                     | 30.570 %HR                 | 20.6 %HR           | -9.970 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.478 %HR                 | 41.8 %HR           | -8.678 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.448 %HR                 | 61.5 %HR           | -8.948 %HR             | ± 5.0 %HR                     | ± 1.9E+00 %HR                        |